=> d his

(FILE 'HOME' ENTERED AT 14:43:48 ON 03 JAN 2005)

FILE 'CA' ENTERED AT 14:44:11 ON 03 JAN 2005

L1 3566 S (NOX OR NO2 OR NITROGEN DIOXIDE) (5A) (CONVERT? OR CONVERSION? OR REDUCTION OR REDUCE# OR REDUCING) (8A) (NITROGEN OXIDE OR NITROGEN MONOXIDE OR NITRIC OXIDE OR "NO")

- L2 649 S L1 AND (DEW OR WATER OR H2O OR MOISTURE OR DEWPOINT)
- L3 308 S L2 AND (FLUE OR STACK OR EXHAUST)
- L4 149 S L2 AND (DETECT? OR DETERMIN? OR ASSAY? OR ANALY? OR ASSESS? OR ESTIMAT? OR EVALUAT? OR MEASUR? OR MONITOR? OR QUANTITAT? OR SENSOR OR SENSING OR SENSE# OR QUANTIF?) (8A) (NOX OR NO2 OR NITROGEN DIOXIDE OR NITROGEN OXIDE OR NITROGEN MONOXIDE OR NITRIC OXIDE OR "NO")
- L5 58 S L3 AND L4
- L6 11 S L3 AND (DRYER OR DRYING OR (WATER OR MOISTURE OR
- H2O) (4A) (ELIMINAT? OR ELLIMINAT? OR REMOV?))
- L7 8 S L3 AND CONDENS?
- L8 91 S L4 NOT L5-7
- L9 16 S L8 AND COMBUST?
- L10 86 S L5-7, L9
- L11 64 S L10 NOT PY>2000
- => d bib, ab 1-64 l11
- L11 ANSWER 1 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 135:246287 CA
- TI Use of a heated sample-inlet, 2-reaction chamber NOx analyzer for combustion optimization
- AU McKarns, Thomas A.; Moser, Werner
- CS ECO Physics, Inc., Ann Arbor, MI, 48108, USA
- SO Proceedings of the Air & Waste Management Association's Annual Conference & Exhibition, 93rd, Salt Lake City, UT, United States, June 18-22, 2000 (2000), 5407-5413 Publisher: Air & Waste Management Association, Pittsburgh, Pa.
- Combustion optimization is key for increasing operating efficiency and for reducing emissions. Higher combustion temps. lead to lower levels of unburned hydrocarbons, but contribute to increased levels of NOx. to fully characterize NOx emissions it is necessary to accurately quantify both NO and NO2 levels. Single reaction chamber NOx analyzers do not correctly report rapidly changing NO2 concns. due to their sequential measurement of NO and NOx levels. The use of sample conditioners (to remove moisture in the sample stream) alters the concn. of NO2 present due to the soly. of NO2 in water. In order to make a correct detn. of the changing concns. of NO and NO2 from a combustion source, a new measuring procedure is By utilizing a chemiluminescence NOx analyzer with heated samplerequired. inlet, maintained at 190 degrees C, the use of a sample conditioner is eliminated for moisture contents up to 15%. Sample integrity is maintained. A two reaction chamber design analyzer, simultaneously measuring NO in one chamber and NOx (after conversion) in the other insures an accurate calcn. of the NO2 concn. during changes in the combustion process. The NO2 concn.

is calcd. by subtracting the NO value from that of the NOx. Single reaction chamber analyzers measure NO first by sending the sample gas directly to the reaction chamber, then, after actuating a solenoid valve, the sample is routed through a converter to the same reaction chamber. The two measurements are made at two different points in time, so if the concns. of NO and NOx are changing with time, the calcn. of NO2 will be incorrect. Addnl., because the solenoid valve disrupts both the sample pressure and flow rate numerous errors are further introduced. With an accurate detn. of the NO and NO2 levels, it is possible to expt. with the combustion conditions to optimize combustion efficiency while concurrently minimizing the circumstances which lead to the formation of higher NOx concns.

- L11 ANSWER 29 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 126:10967 CA
- TI In-situ FT-IR measurements at industrial combustion facilities
- AU Bonanno, Anthony S.; Nelson, Chad M.; Wojtowicz, Marek A.; Knight, Kim S.; Serio, Michael A.; Solomon, Peter R.
- CS Advanced Fuel Research, Inc., East Hartford, CT, 06108, USA
- SO Proceedings of SPIE-The International Society for Optical Engineering (1996), 2883 (Optical Remote Sensing for Environmental and Process Monitoring), 611-616
- AB This paper describes the development and testing of a prototype FTIR-based measurement system for continuous emission monitoring (CEM) and process control in fossil fuel-fired power plants. On several occasions, prototype systems have been transported and assembled at full-scale and pilot-scale fossil fuel-fired combustors. The in-situ version of the prototype is capable of measuring NH3 and HCl concns., which are difficult to measure extractively, as well as CO, CO2, NOx, H2O, and SOx concns. The results are presented of recent tests which involve in-situ monitoring of selective noncatalytic redn. (SNCR) of NOx based on simultaneous measurement of NO, NH3, and CO.
- L11 ANSWER 42 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 110:12726 CA
- TI Measurement of nitrogen oxides in ambient air and in stack gas. (3).
- AU Tamaki, Motonori; Hiraki, Takatoshi
- CS Hyogokenritsu Kogai Kenkyusho, Japan
- SO Kankyo Gijutsu (1988), 17(6), 396-400
- LA Japanese
- AB The NOx concn. in air and from stationary sources was detd. by the chemiluminescence method, and the problems assocd. with the method are discussed. In the method, NO2 was catalytically reduced to NO by a catalytic converter. The selectivity of NOx is less than desirable; there is NH3 conversion to NO and interference by water.
- L11 ANSWER 46 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 95:191704 CA
- TI An apparatus for determination of ammonia in flue gases
- PA Seitetsu Kagaku Co., Ltd., Japan
- SO Jpn. Tokkyo Koho, 5 pp.
- PI JP 56032580 B4 19810729 JP 1977-3167 19770114

PRAI JP 1977-3167 A 19770114

AB NH3 in **flue** gases (contg. **NOx**, SOx, O, NH3, and **moisture**) is **monitored** by a chemiluminescence-type app. designed to convert NH3 to **NO** for **anal**. without forming NH4HSO3. The SOx is removed by sorption with soda lime, and the SOx-free **flue** gas is divided into 2 streams. One stream passes through a Cu-Cr series catalyst at 300-450° to **convert** NH3 to **NO2** and **NO**, and then through a com. **redn**. catalyst to **reduce NO2** to **NO**. This stream and the other one are analyzed and the results are compared.

- L11 ANSWER 48 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 91:95830 CA
- TI Analyzers for **flue** gases from incinerators
- AU Kaneko, Teruo
- CS Fuji Electr. Co., Hino, Japan
- SO Fuji Jiho (1978), 51(12), 644-8
- LA Japanese

AB Analyzers for HCl, NOx, SO2, CO, O, and CO2 flue gas from incinerators are developed. HCl 5-2000 ppm is collected into KNO3 soln. and detd. by ion electrode. NO and SO2 0-1000 ppm are detd. by a nondispersive IR analyzer with flow rate of 2.5 L/min. NO2 is converted to NO. O 0-25% is detd. by a magnetic wind-type analyzer. CO and CO2 are also detd. by the nondispersive IR analyzer. To prevent the loss of water-sol. gases, a gas conducting tube between the stack and the analyzer is heated >120°. Fine dust particles are removed by membrane filter which is located upstream of the analyzer.

- L11 ANSWER 50 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 90:209265 CA
- TI Improved method for **determining** the concentration of **nitrogen oxides** in **exhaust** gases
- AU Yokoyama, Kimihiko; Kira, Naruto; Watanabe, Tetsuya; Uchiyama, Haruo
- CS Tech. Res. Lab., Ishikawajima-Harima Heavy Ind. Co., Ltd., Tokyo, Japan
- SO Ishikawajima-Harima Giho (1977), 17(6), 574-8
- LA Japanese
- AB In detg. the concn. of NOx, i.e. NO + NO2, in flue gases from boilers by the chemiluminescent method, the gas was contacted at the sampling port with activated C at 240-300° to reduce NO2 to NO, thus avoiding errors due to absorption of NO2 by water condensed in the sampling line and the moisture eliminator. The concn. of NO was detd. after removing NO2 at the sampling port by contact with aq. soln. (pH 5-7) of Na2SO3 (10 wt.%) contg. CuCl2 (0.1M) and NaCl (3 wt.%).
- L11 ANSWER 57 OF 64 CA COPYRIGHT 2005 ACS on STN
- AN 87:10596 CA
- TI Some considerations in determining oxides of nitrogen in **stack** gases by chemiluminescence analyzer
- AU Heyman, G. A.; Turner, G. S.
- CS Beckman Instrum., Fullerton, CA, USA
- SO ISA Transactions (1976), 15(3), 209-13
- AB Federal regulations require the continuous monitoring of NOx in emissions from fossil fuel-fired steam generators and HNO3 plants. Chemiluminescence analyzers are being used for this purpose. These

analyzers contain a reaction chamber for the reaction of O3 and NOx, a sensitive photomultiplier tube used to detect the radiation produced in this reaction, a source of O3 and a NO2-to-NO converter. Sample handling requirements are specific for each installation. Requirements for the total anal. system are discussed. Wet and dry basis anal. are considered and potential interference from CO, H2O, NH3 and SO2 are evaluated.

=> log y STN INTERNATIONAL LOGOFF AT 15:12:25 ON 03 JAN 2005